

240 County Road Ipswich, MA 01938-2723 Tel 978-927-5054 Fax 978-921-1350 www.neb.com info@neb.com

## New England Biolabs Certificate of Analysis

Product Name: NEBNext® rRNA Depletion Kit v2 (Human/Mouse/Rat)

Catalog Number: E7400X
Packaging Lot Number: 10147793
Expiration Date: 08/2023
Storage Temperature: -20°C

Specification Version: PS-E7400S/L/X v1.0

NEBNext® rRNA Depletion Kit v2 (Human/Mouse/Rat) Component List				
NEB Part Number	Component Description	Lot Number	Individual QC Result	
E7753-4VIAL	NEBNext® DNase I	10116536	Pass	
E7752-4VIAL	NEBNext® Thermostable RNase H	10116537	Pass	
E7401-4VIAL	NEBNext® V2 rRNA Depletion Solution	10116538	Pass	
E6317-4VIAL	Nuclease-free Water	10116541	Pass	
E6315-4VIAL	DNase I Reaction Buffer	10116540	Pass	
E6314-4VIAL	NEBNext® Probe Hybridization Buffer	10116542	Pass	
E6312-4VIAL	RNase H Reaction Buffer	10116539	Pass	

Assay Name/Specification	Lot # 10147793
Functional Testing (rRNA Depletion) NEBNext® rRNA Depletion Kit v2 (Human/Mouse/Rat) is functionally validated using commercially available human RNA (e.g. UHRR). After treatment with the kit, a high-throughput sequencing library is made and sequenced. Reads from this library are classified as being derived from ribosomal RNA based on sequence homology. This method produces libraries with <10% rRNA.	Pass
* Individual Product Component Note Standard Quality Control Tests are performed for each component included in NEBNext® rRNA Depletion Kit v2 (Human/Mouse/Rat) and meet the designated specifications.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



E7400X / Lot: 10147793 Page 1 of 2 Christin Summ

Christine Sumner Production Scientist 06 Apr 2022 Michael Tonello

Packaging Quality Control Inspector

06 Apr 2022

E7400X / Lot: 10147793

Page 2 of 2