

## New England Biolabs Certificate of Analysis

Product Name: NEBNext<sup>®</sup> RNA Depletion Core Reagent Set  
 Catalog Number: E7865S  
 Packaging Lot Number: 10180713  
 Expiration Date: 04/2025  
 Storage Temperature: -20°C  
 Specification Version: PS-E7865S/L/X v1.0

NEBNext <sup>®</sup> RNA Depletion Core Reagent Set Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
E7753-2VIAL	NEBNext <sup>®</sup> DNase I	10180714	Pass
E7752-2VIAL	NEBNext <sup>®</sup> Thermostable RNase H	10180715	Pass
E6317-2VIAL	Nuclease-free Water	10180718	Pass
E6315-2VIAL	DNase I Reaction Buffer	10180717	Pass
E6314-2VIAL	NEBNext <sup>®</sup> Probe Hybridization Buffer	10180719	Pass
E6312-2VIAL	RNase H Reaction Buffer	10180716	Pass

Assay Name/Specification	Lot # 10180713
<p><b>* Individual Product Component Note</b>            Standard Quality Control Tests are performed for each component included in NEBNext<sup>®</sup> RNA Depletion Core Reagent Set and meet the designated specifications.</p>	Pass
<p><b>Functional Testing (rRNA Depletion)</b>            NEBNext<sup>®</sup> RNA Depletion Core Reagent Set is functionally validated using commercially available human RNA (e.g. UHRR). After treatment with the kit, a high-throughput sequencing library is made and sequenced. Reads from this library are classified as being derived from ribosomal RNA based on sequence homology. This method produces libraries with &lt;10% rRNA.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.

*Christine Sumner*

---

Christine Sumner  
Production Scientist  
03 May 2023

*Michael Tonello*

---

Michael Tonello  
Packaging Quality Control Inspector  
16 Jan 2024