

## New England Biolabs Certificate of Analysis

**Product Name:** AlwNI  
**Catalog Number:** R0514L  
**Concentration:** 10,000 U/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.  
**Packaging Lot Number:** 10133247  
**Expiration Date:** 12/2022  
**Storage Temperature:** -20°C  
**Storage Conditions:** 50 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA  
**Specification Version:** PS-R0514S/L v1.0

AlwNI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0514LVIAL	AlwNI	10092949	Pass
B6004SVIAL	rCutSmart™ Buffer	10130601	Pass

Assay Name/Specification	Lot # 10133247
<b>Ligation and Recutting (Terminal Integrity)</b> After a 20-fold over-digestion of Lambda DNA with AlwNI, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with AlwNI.	Pass
<b>Non-Specific DNase Activity (16 Hour)</b> A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 10 Units of AlwNI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
<b>Exonuclease Activity (Radioactivity Release)</b> A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [ <sup>3</sup> H] E. coli DNA and a minimum of 100 units of AlwNI incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
<b>Protein Purity Assay (SDS-PAGE)</b> AlwNI is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit [www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Pengda Zhang  
Production Scientist  
19 Jan 2022



Josh Hersey  
Packaging Quality Control Inspector  
19 Jan 2022