

## New England Biolabs Certificate of Analysis

**Product Name:** Hpy166II  
**Catalog #:** R0616S/L  
**Concentration:** 10,000 units/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1  $\mu$ g of pBR322 in 1 hour at 37°C in total reaction volume of 50  $\mu$ l.  
**Lot #:** 0101604  
**Assay Date:** 04/2016  
**Expiration Date:** 4/2018  
**Storage Temp:** -20°C  
**Storage Conditions:** 250 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 0.15% Triton X-100, 200  $\mu$ g/ml BSA  
**Specification Version:** PS-R0616S/L v1.0  
**Effective Date:** 19 Feb 2015

Assay Name/Specification (minimum release criteria)	Lot #0101604
<b>Exonuclease Activity (Radioactivity Release)</b> - A 50 $\mu$ l reaction in CutSmart™ Buffer containing 1 $\mu$ g of a mixture of single and double-stranded [ <sup>3</sup> H] <i>E. coli</i> DNA and a minimum of 30 units of Hpy166II incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
<b>Ligation and Recutting (Terminal Integrity)</b> - After a 10-fold over-digestion of pBR322 DNA with Hpy166II, >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Hpy166II.	Pass
<b>Non-Specific DNase Activity (16 Hour)</b> - A 50 $\mu$ l reaction in CutSmart™ Buffer containing 1 $\mu$ g of pBR322 DNA and a minimum of 50 units of Hpy166II incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
<b>Protein Purity Assay (SDS-PAGE)</b> - Hpy166II is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.	Pass

\* The BSA in this product has been granted an EDQM "Certificate of Suitability" from the European Directorate for the Quality of Medicines (# R1-CEP-2003-204-Rev00) and has been granted a USDA Certificate for Export of Bovine Blood Plasma/Serum for Manufacture into Pharmaceutical Products.



Authorized by  
Derek Robinson  
19 Feb 2015



Inspected by  
Stephanie Cornelio  
22 Mar 2016

