

## New England Biolabs Certificate of Analysis

**Product Name:** PspXI  
**Catalog Number:** R0656S  
**Concentration:** 5,000 U/ml  
**Unit Definition:** One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA (HindIII digest) in 1 hour at 37°C in a total reaction volume of 50 µl.  
**Packaging Lot Number:** 10091961  
**Expiration Date:** 12/2022  
**Storage Temperature:** -20°C  
**Storage Conditions:** 300 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 500 µg/ml BSA  
**Specification Version:** PS-R0656S/L v1.0

PspXI Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R0656SVIAL	PspXI	10091962	Pass
B7204SVIAL	CutSmart® Buffer	10091032	Pass

Assay Name/Specification	Lot # 10091961
<p><b>Non-Specific DNase Activity (16 Hour)</b>            A 50 µl reaction in CutSmart™ Buffer containing 1 µg of Lambda-HindIII DNA and a minimum of 25 Units of PspXI incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	Pass
<p><b>Exonuclease Activity (Radioactivity Release)</b>            A 50 µl reaction in CutSmart™ Buffer containing 1 µg of a mixture of single and double-stranded [<sup>3</sup>H] E. coli DNA and a minimum of 50 units of PspXI incubated for 4 hours at 37°C releases &lt;0.1% of the total radioactivity.</p>	Pass
<p><b>Ligation and Recutting (Terminal Integrity)</b>            After a 10-fold over-digestion of Lambda-HindIII DNA with PspXI, &gt;95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, &gt;95% can be recut with PspXI.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit

[www.neb.com/trademarks](http://www.neb.com/trademarks) for additional information.



Penghua Zhang  
Production Scientist  
18 Dec 2020



Michael Tonello  
Packaging Quality Control Inspector  
18 Dec 2020