

New England Biolabs Certificate of Analysis

Product Name: Scal-HF[®]
Catalog Number: R3122S
Concentration: 20,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.
Packaging Lot Number: 10089670
Expiration Date: 01/2022
Storage Temperature: -20°C
Storage Conditions: 200 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R3122S/L v1.0

Scal-HF [®] Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R3122SVIAL	Scal-HF [®]	10064291	Pass
B7204SVIAL	CutSmart [®] Buffer	10085424	Pass
B7024AVIAL	Gel Loading Dye, Purple (6X)	10082935	Pass

Assay Name/Specification	Lot # 10089670
Endonuclease Activity (Nicking) A 50 µl reaction in CutSmart [™] Buffer containing 1 µg of supercoiled PhiX174 DNA and a minimum of 20 Units of Scal-HF [™] incubated for 4 hours at 37°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart [™] Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 100 units of Scal-HF [™] incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Ligation and Recutting (Terminal Integrity) After a 2-fold over-digestion of Lambda DNA with Scal-HF [™] , >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Scal-HF [™] .	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart [™] Buffer containing 1 µg of Lambda DNA and a minimum of 60 Units of Scal-HF [™] incubated for 16 hours at 37°C results in a DNA pattern free of	Pass

Assay Name/Specification	Lot # 10089670
<p>detectable nuclease degradation as determined by agarose gel electrophoresis.</p> <p>Protein Purity Assay (SDS-PAGE) Scal-HF™ is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.</p>	<p>Pass</p>

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



Penghua Zhang
Production Scientist
02 Nov 2020



Josh Hersey
Packaging Quality Control Inspector
02 Nov 2020