

New England Biolabs Certificate of Analysis

Product Name: Sspl-HF[®]
Catalog Number: R3132M
Concentration: 100,000 U/ml
Unit Definition: One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in 1 hour at 37°C in a total reaction volume of 50 µl.
Packaging Lot Number: 10061991
Expiration Date: 11/2021
Storage Temperature: -20°C
Storage Conditions: 200 mM NaCl, 10 mM Tris-HCl (pH 7.4), 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml BSA
Specification Version: PS-R3132M v1.0

Sspl-HF [®] Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R3132M VIAL	Sspl-HF [®]	10058512	Pass
B7204S VIAL	CutSmart [®] Buffer	10061299	Pass
B7024S VIAL	Gel Loading Dye, Purple (6X)	10055732	Pass

Assay Name/Specification	Lot # 10061991
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in CutSmart [™] Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 100 units of Sspl-HF [™] incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Non-Specific DNase Activity (16 Hour) A 50 µl reaction in CutSmart [™] Buffer containing 1 µg of Lambda DNA and a minimum of 100 Units of Sspl-HF [™] incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.	Pass
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with Sspl-HF [™] , >95% of the DNA fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with Sspl-HF [™] .	Pass

This product has been tested and shown to be in compliance with all specifications.



Anthony Francis
Production Scientist
02 Dec 2019



Jay Minichiello
Packaging Quality Control Inspector
08 Jan 2020