

New England Biolabs Certificate of Analysis

Product Name: *Sbfl-HF[®]*
Catalog Number: *R3642S*
Concentration: *20,000 U/ml*
Unit Definition: *One unit is defined as the amount of enzyme required to digest 1 µg of Lambda DNA in rCutSmart Buffer in 1 hour at 37°C in a total reaction volume of 50 µl.*
Packaging Lot Number: *10187713*
Expiration Date: *05/2025*
Storage Temperature: *-20°C*
Storage Conditions: *10 mM Tris-HCl, 200 mM NaCl, 1 mM DTT, 0.1 mM EDTA, 50% Glycerol, 200 µg/ml rAlbumin (pH 7.4 @ 25°C)*
Specification Version: *PS-R3642S/L v2.0*


Sbfl-HF [®] Component List			
NEB Part Number	Component Description	Lot Number	Individual QC Result
R3642SVIAL	Sbfl-HF [®]	10187676	Pass
B7024AVIAL	Gel Loading Dye, Purple (6X)	10182168	Pass
B6004SVIAL	rCutSmart [™] Buffer	10182170	Pass

Assay Name/Specification	Lot # 10187713
Endonuclease Activity (Nicking) A 50 µl reaction in rCutSmart [™] Buffer containing 1 µg of supercoiled pBR322 DNA and a minimum of 20 units of Sbfl-HF [®] incubated for 4 hours at 37°C results in <20% conversion to the nicked form as determined by agarose gel electrophoresis.	Pass
Exonuclease Activity (Radioactivity Release) A 50 µl reaction in rCutSmart [™] Buffer containing 1 µg of a mixture of single and double-stranded [³ H] E. coli DNA and a minimum of 100 units of Sbfl-HF [®] incubated for 4 hours at 37°C releases <0.1% of the total radioactivity.	Pass
Functional Testing (15 minute Digest) A 50 µl reaction in rCutSmart [™] Buffer containing 1 µg of Lambda DNA and 1 µl of Sbfl-HF [®] incubated for 15 minutes at 37°C results in complete digestion as determined by agarose gel electrophoresis.	Pass
Ligation and Recutting (Terminal Integrity) After a 10-fold over-digestion of Lambda DNA with Sbfl-HF [®] , >95% of the DNA	Pass


Assay Name/Specification	Lot # 10187713
fragments can be ligated with T4 DNA ligase in 16 hours at 16°C. Of these ligated fragments, >95% can be recut with SbfI-HF®.	
<p>Non-Specific DNase Activity (16 Hour) A 50 µl reaction in rCutSmart™ Buffer containing 1 µg of Lambda DNA and a minimum of 20 units of SbfI-HF® incubated for 16 hours at 37°C results in a DNA pattern free of detectable nuclease degradation as determined by agarose gel electrophoresis.</p>	Pass
<p>Protein Purity Assay (SDS-PAGE) SbfI-HF® is >95% pure as determined by SDS PAGE analysis using Coomassie Blue detection.</p>	Pass

This product has been tested and shown to be in compliance with all specifications.

One or more products referenced in this document may be covered by a 3rd-party trademark. Please visit www.neb.com/trademarks for additional information.



YunJie Sun
Production Scientist
02 May 2023



Michael Tonello
Packaging Quality Control Inspector
05 May 2023